Datasheet

Version: 1.0.0 Revision date: 25 Jun 2025



Human Regulator of G-protein signaling 21 (RGS21) ELISA Kit

Catalogue No.:abx542135

Human Regulator of G-protein signaling 21 (RGS21) ELISA Kit is an ELISA Kit for the in vitro quantitative measurement of Human Regulator of G-protein signaling 21 (RGS21) concentrations in serum, plasma, tissue homogenates, cell culture supernatants and other biological fluids.

Target: Regulator of G-protein signaling 21 (RGS21)

Reactivity: Human

Tested Applications: ELISA

Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.

Storage: Shipped at 4 °C. Upon receipt, store the kit according to the storage instruction in the kit's manual.

Validity: The validity for this kit is 6 months.

Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within

the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested

that the whole assay is performed by the same user throughout.

UniProt Primary AC: Q2M5E4 (<u>UniProt</u>, <u>ExPASy</u>)

Gene Symbol: RGS21

GenelD: <u>431704</u>

OMIM: <u>612407</u>

HGNC: 26839

KEGG: hsa:431704

Test Range: 0.156 ng/ml - 10 ng/ml

Standard Form: Lyophilized

Detection Method: Colorimetric

Assay Data: Quantitative

Datasheet

Version: 1.0.0 Revision date: 25 Jun 2025



Sample Type: Serum, plasma, tissue homogenates, cell culture supernatants and other biological fluids.

Note: THIS PRODUCT IS FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC OR

THERAPEUTIC PROCEDURES.

The range and sensitivity is subject to change. Please contact us for the latest product information. For accurate results, sample concentrations must be diluted to mid-range of the kit. If you require a specific range, please contact us in advance or write your request in your order comments. Please note that our kits are optimised for detection of native samples, rather than recombinant proteins. We are unable to guarantee detection of recombinant proteins, as they may have different sequences or tertiary structures to the native protein.