

Mouse Estrogen Receptor Beta (ESR2) CLIA Kit

Catalogue No.: abx195591

Mouse Estrogen Receptor 2 Chemiluminescent Immunoassay (CLIA) Kit is a Sandwich Chemiluminescent Immunoassay (CLIA) Kit for use with Tissue homogenates, cell lysates and other biological fluids.

Target:	Estrogen Receptor Beta (ESR2)
Research Area:	Endocrinology, Reproductive Science, Hormone Metabolism
Reactivity:	Mouse
Tested Applications:	CLIA
Recommended dilutions:	Optimal dilutions/concentrations should be determined by the end user.
Storage:	Shipped at 4°C. Upon receipt, store the kit according to the storage instruction in the kit's manual.
Validity:	The validity for this kit is at least 6 months. Up to 12 months validity can be provided on request.
Stability:	The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.
UniProt Primary AC:	O08537 (UniProt , ExPASy)
Gene Symbol:	ESR2
GeneID:	13983
KEGG:	mmu:13983
String:	10090.ENSMUSP00000098849
Test Range:	0.312 ng/ml - 20 ng/ml
Sensitivity:	< 0.142 ng/ml
Standard Form:	Lyophilized
Detection Method:	Chemiluminescent

Datasheet

Version: 3.0.0

Revision date: 06 Oct 2025



Assay Type: Sandwich

Assay Data: Quantitative

Sample Type: Tissue homogenates, cell lysates and other biological fluids.

Note: THIS PRODUCT IS FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC OR THERAPEUTIC PROCEDURES.

The range and sensitivity is subject to change. Please contact us for the latest product information.

For accurate results, sample concentrations must be diluted to mid-range of the kit. If you require a specific range, please contact us in advance or write your request in your order comments.

Please note that our kits are optimised for detection of native samples, rather than recombinant proteins. We are unable to guarantee detection of recombinant proteins, as they may have different sequences or tertiary structures to the native protein.

For Reference Only