

## Glutamate Decarboxylase 2 (GAD2) Antibody Pair

Catalogue No.: abx370729

Glutamate Decarboxylase 2 (GAD2) Antibody Pair for use in Sandwich ELISA assay development. This antibody pair contains:

| Component                            | 5 × 96 tests | 10 × 96 tests |
|--------------------------------------|--------------|---------------|
| Capture Antibody                     | 200 µg       | 400 µg        |
| Biotin-Conjugated Detection Antibody | 50 µg        | 100 µg        |
| Standard                             | 2 µg         | 10 µg         |

Please note that quantities and concentrations may change between different batches.

It is recommended to use this antibody pair with [abx098958 Antibody Pair Support Kit \(Sandwich Method\)](#).

|                               |  |
|-------------------------------|--|
| <b>Target:</b>                | Glutamate Decarboxylase 2 (GAD2)   |
| <b>Reactivity:</b>            | Rat  |
| <b>Tested Applications:</b>   | ELISA  |
| <b>Recommended dilutions:</b> | Dilute the Capture Antibody 125-fold with Coating Buffer.<br>Dilute the Biotin-Conjugated Detection Antibody 200-fold with Detection Antibody Diluent.<br>Optimal dilutions/concentrations should be determined by the end user. |
| <b>Form:</b>                  | Liquid (Capture Antibody and Detection Antibody)   |
| <b>Reconstitution:</b>        | Reconstitute the standard with Standard Diluent. The volume, and therefore standard concentration, should be determined by the end user.   |
| <b>Storage:</b>               | Aliquot and store at -20°C. Avoid repeated freeze/thaw cycles.   |
| <b>UniProt Primary AC:</b>    | Q05683 ( <a href="#">UniProt</a> , <a href="#">ExPASy</a> )  |
| <b>Gene Symbol:</b>           | GAD2   |
| <b>GeneID:</b>                | <a href="#">24380</a>  |
| <b>KEGG:</b>                  | rno:24380  |
| <b>Ensembl:</b>               | ENSRNOG00000018200   |
| <b>String:</b>                | <a href="#">10116.ENSARNOP00000024901</a>  |

# Datasheet

Version: 3.0.0  
Revision date: 19 Jul 2025



|  |  |
|--|--|
| <b>Buffer:</b>                         | The Capture and Detection Antibody both contain 0.1% sodium azide.   |
| <b>Standard Form:</b>                  | Lyophilized  |
| <b>Assay Type:</b>                     | Sandwich   |
| <b>Capture Antibody Conjugation:</b>   | Unconjugated   |
| <b>Detection Antibody Conjugation:</b> | Biotin   |
| <b>Concentration:</b>                  | Capture Antibody: 0.5 mg/ml<br>Biotin-Conjugated Detection Antibody: 0.2 mg/ml   |
| <b>Note:</b>                           | THIS PRODUCT IS FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC, THERAPEUTIC OR COSMETIC PROCEDURES. NOT FOR HUMAN OR ANIMAL CONSUMPTION.   |
| <b>Directions for use:</b>             | <p>Bring all components to room temperature (18-25°C) and briefly spin or centrifuge the vials before use. Working solutions should be prepared and used immediately.</p> <p><u>Recommended Procedure:</u></p> <ol style="list-style-type: none"><li>1. Dilute the Capture Antibody to working concentration using Coating Buffer. Immediately coat the 96-well plate with diluted Capture Antibody (100 µl per well). Seal the plate and incubate at 4 °C overnight or at 37 °C for 2 hours</li><li>2. Aspirate the wells and wash with Wash Buffer (350 µl per well) and allow to soak for 1-2 min. Remove the liquid by inverting and tapping the plate on to absorbent paper.</li><li>3. Block the plate with Blocking Buffer (200 µl per well) at 37 °C for 1.5 hours.</li><li>4. Repeat the aspiration/wash process in Step 2.</li><li>5. Add 100 µl of standards or sample into the appropriate wells. Cover with a plate sealer and incubate at 37 °C for 1 hour.</li><li>6. Repeat the aspiration/wash process in Step 2.</li><li>7. Add appropriately diluted Biotin-Conjugated Detection Antibody (100 µl per well). Cover the plate with a new plate sealer and incubate at 37 °C for 1 hour.</li><li>8. Repeat the aspiration/wash process in Step 2.</li><li>9. Add appropriately diluted Streptavidin HRP (100 µl per well). Cover the plate with a new plate sealer and incubate at 37 °C for 30 min.</li><li>10. Repeat the aspiration/wash process in Step 2.</li><li>11. Add Substrate Solution (90 µl per well). Cover the plate with a new plate sealer and incubate at 37 °C for 10-20 min. Keep the plate in the dark and avoid exposure to light.</li><li>12. Add Stop Solution (50 µl per well). Tap the side of the plate to ensure thorough mixing.</li><li>13. Measure the absorbance immediately using a microplate reader set at 450 nm.</li></ol> |