Version: 6.0.0 Revision date: 17 Oct 2025



Human Serum Amyloid A-1 Protein (SAA1) ELISA Kit

Catalogue No.:abx153065

Human Serum Amyloid A-1 Protein (SAA1) ELISA Kit is an ELISA Kit for the in vitro quantitative measurement of Human Serum Amyloid A (SAA1) concentrations in serum, plasma, tissue homogenates, cell lysates, cell culture supernatants and other biological fluids. This assay has high sensitivity and excellent specificity for detection of Serum Amyloid A (SAA1) No significant cross-reactivity or interference between Serum Amyloid A (SAA1) and analogues was observed.

Target: Serum Amyloid A-1 Protein (SAA1)

Research Area: Infection Immunity, Kidney Biomarkers, Hepatology

Reactivity: Human

Tested Applications: ELISA

Recommended Optimal dilutions/concentrations should be determined by the end user.

dilutions:

Storage: Shipped at 4°C. Upon receipt, store the kit according to the storage instruction in the kit's manual.

Validity: The validity for this kit is at least 6 months. Up to 12 months validity can be provided on request.

Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within

the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested

that the whole assay is performed by the same user throughout.

UniProt Primary AC: P0DJI8 (UniProt, ExPASy)

Gene Symbol: SAA1

GeneID: 6288

OMIM: <u>104750</u>

KEGG: hsa:6288

Test Range: 625 ng/ml - 10000 ng/ml

Sensitivity: < 251 ng/ml

Standard Form: Lyophilized

Detection Method: Colorimetric

Version: 6.0.0 Revision date: 17 Oct 2025



Assay Type: Competitive

Assay Data: Quantitative

Sample Type: Serum, plasma, tissue homogenates, cell lysates, cell culture supernatants and other biological

fluids.

Target Type: Antigen

Assay Principle: This kit is based on competitive enzyme-linked immuno-sorbent assay technology. An antibody is

pre-coated onto a 96-well plate. Standards, test samples, and biotin-conjugated reagent are added to the wells and incubated. A competitive inhibition reaction takes place between the biotin-labelled SAA1 and the unlabelled- SAA1 on the pre-coated antibody. The HRP-conjugated reagent is then added, and the whole plate is incubated. Unbound conjugates are removed using wash buffer at each stage. TMB substrate is used to quantify the HRP enzymatic reaction. After TMB substrate is added, only wells that contain sufficient SAA1 will produce a blue coloured product, which then changes to yellow after adding the acidic stop solution. The intensity of the color yellow is inversely proportional to the SAA1 amount bound on the plate. The OD is measured spectrophotometrically at 450 nm in a microplate reader, from which the concentration of SAA1 can be calculated.

450 nm in a micropiate reader, from which the concentration of SAA1 can be calculated.

The kit components listed are for reference only. The product manual may differ slightly. The product should be used as stated on the product manual included and delivered together with the

product.

Pre-coated 96-Well Microplate

Standard

Standard Diluent Buffer

Wash Buffer

Detection Reagent A

Detection Reagent B

• Diluent A

• Diluent B

TMB Substrate

Stop Solution

Plate Sealer

Material Required But

• 37°C incubator

Not Provided:

Kit Components:

· Multi and single channel pipettes and sterile pipette tips

• Squirt bottle or automated microplate washer

• 1.5 ml tubes

Distilled water

· Absorbent filter papers

• 100 ml and 1 liter graduated cylinders

• Microplate reader (wavelength: 450 nm)

• ELISA Shaker

Version: 6.0.0 Revision date: 17 Oct 2025



Reagent Preparation:

This procedure is provided for reference only. The product manual may differ slightly. The product should be used as stated on the product manual included and delivered together with the product.

- 1) Standard: Prepare the standard with the recommended volume of Standard Diluent Buffer, to make the standard solution. Then use the Standard Diluent buffer to carry out serial dilutions of the standard solution, as instructed in the Protocol.
- 2) Wash Buffer: Dilute the concentrated Wash Buffer with distilled water, as instructed in the Protocol.
- 3) Detection Reagent Preparation: Calculate the total volume of working solution required. Dilute Detection Reagent A and Detection Reagent B with Diluent A and Diluent B, respectively, at 1:100.

Assay Procedure:

This procedure is provided for reference only. The product manual may differ slightly. The product should be used as stated on the product manual included and delivered together with the product.

- 1) Set standard, test samples and control wells.
- 2) Aliquot 50 µl of diluted standard into the standard wells.
- 3) Aliquot 50 µl of Standard Diluent buffer into the control (zero) well.
- 4) Aliquot 50 µl of diluted samples into the sample wells.
- 5) Immediately aliquot 50 µl of Detection Reagent A to each well. Incubate for 1 hr at 37 °C.
- 6) Wash 3 times.
- 7) Aliquot 100 µl of Detection Reagent B to each well. Incubate for 30 mins at 37 °C.
- 8) Wash 5 times.
- 9) Aliquot 90 µl of TMB Substrate to each well. Incubate for 10-20 mins at 37 °C.
- 10) Aliquot 50 µl of Stop Solution.
- 11) Measure the OD at 450 nm.

Protocol:

This procedure is provided for reference only. The product manual may differ slightly. The product should be used as stated on the product manual included and delivered together with the product.

- Equilibrate the kit components and samples to room temperature (18 25 °C) before use. It is recommended to plot a standard curve for each test.
- 1. Set standard, test sample and control (zero) wells on the pre-coated plate respectively, and then, record their positions. It is recommended to measure each standard and sample at least in duplicate.
- 2. Add 50 µL of each standard, control and sample into the appropriate wells.
- 3. Remove the cover and discard the liquid.
- 4. Immediately aliquot 50 μl of Detection Reagent A working solution. Seal the plate with a cover and incubate for 1 h at 37°C.
- 5. Remove the cover and discard the solution. Wash the plate 3 times with 1X Wash Buffer.
- \bullet 6. Add 100 μL of Detection Reagent B working solution into each well, seal and incubate at 37°C for 30 min.
- 7. Discard the solution and wash the plate 5 times with wash buffer as explained in previous step.
- 8. Aliquot 90 µl of TMB Substrate into each well. Seal the plate with a cover and incubate at 37°C for 10-20 min. Avoid exposure to light. The incubation time is for reference use only, the optimal time should be determined by end user. Do not exceed 30 min.
- 9. Add 50 µL of Stop Solution to each well. Read at 450 nm immediately.

Version: 6.0.0 Revision date: 17 Oct 2025



Results Calculation: This assay is competitive, therefore there is an inverse correlation between SAA1 concentration in

the sample and the absorbance measured. Create a graph with the log of the standard

concentration (y-axis) and average absorbance measured (x-axis). Apply a best fit trendline through the standard points. The SAA1 concentration of the samples can be interpolated from the standard

curve.

Assay Precision: Intra-assay Precision (Precision within an assay): 3 samples with low, medium and high levels of

Serum Amyloid A (SAA) were were tested 20 times on one plate, respectively.

Inter-assay Precision (Precision between assays): 3 samples with low, medium and high levels of

Serum Amyloid A (SAA) were tested on 3 different plates, 8 replicates in each plate.

CV (%) = (Standard Deviation / mean) × 100

Intra-Assay: CV<10% Inter-Assay: CV<12%

Note: THIS PRODUCT IS FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC OR

THERAPEUTIC PROCEDURES.

The range and sensitivity is subject to change. Please contact us for the latest product information. For accurate results, sample concentrations must be diluted to mid-range of the kit. If you require a specific range, please contact us in advance or write your request in your order comments. Please note that our kits are optimised for detection of native samples, rather than recombinant

proteins. We are unable to guarantee detection of recombinant proteins, as they may have different

sequences or tertiary structures to the native protein.

